



Arthrex Inc.
Ivette Galmez
Official Correspondent, Regulatory Affairs
1370 Creekside Boulevard
Naples, Florida 34108-1945

June 7, 2021

Re: K210050

Trade/Device Name: Augmented VaultLock Glenoid
Regulation Number: 21 CFR 888.3660
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis
Regulatory Class: Class II
Product Code: KWS
Dated: May 21, 2021
Received: May 25, 2021

Dear Ivette Galmez:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part

801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Michael C. Owens -S Digitally signed by
Michael C. Owens -S
Date: 2021.06.07
17:40:04 -04'00'

Michael Owens
Assistant Director
DHT6A: Division of Joint
Arthroplasty Devices
OHT6: Office of Orthopedic Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K210050

Device Name

Augmented VaultLock Glenoid

Indications for Use (Describe)

The Augmented VaultLock Glenoid is indicated in replacement(s) when conditions include severe pain or significant disability resulting from degenerative, rheumatoid, traumatic disease, or injury of the glenohumeral joint; non-union humeral head fractures of long duration; irreducible 2- and 4- part proximal humeral fractures; avascular necrosis of the humeral head; or, other difficult clinical management problems where arthrodesis or resectional arthroplasty is not acceptable.

The glenoid components are designed for cemented fixation in the joint and must only be used with appropriate bone cement.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

Date Prepared	December 24, 2020
Submitter	Arthrex Inc. 1370 Creekside Boulevard Naples, FL 34108-1945
Contact Person	Ivette Galmez Regulatory Affairs 1-239-643-5553, ext. 71263 ivette.galmez@arthrex.com
Name of Device	Augmented VaultLock Glenoid
Common Name	Shoulder Prosthesis
Product Code	KWS
Classification Name	21 CFR 888.3660: Prosthesis, Shoulder, semi-constrained metal/polymer, cemented
Regulatory Class	II
Predicate Device	K161108: Arthrex VaultLock Glenoid
Reference Device	K191960: Arthrex Univers Revers Modular Glenoid System K071032: Arthrex Univers II Shoulder Prosthesis
Purpose of Submission	This Traditional 510(k) premarket notification is submitted to obtain clearance for the Augmented VaultLock Glenoid for use with the existing Univers II Shoulder Prosthesis system (K071032).
Device Description	The Augmented VaultLock Glenoid is made of the same materials as the predicate (UHMWPE). The Augmented VaultLock Glenoid is designed with a half-wedge augment. The proposed device has an identical spherical articulating surface as that of the previously cleared glenoids and is available in 4 nominal sizes. The proposed device is a line extension to the Arthrex VaultLock Glenoid cleared under K161108.
Indications for Use	The Augmented VaultLock Glenoid is indicated in replacement(s) when conditions include severe pain or significant disability resulting from degenerative, rheumatoid, traumatic disease, or injury of the glenohumeral joint; non-union humeral head fractures of long duration; irreducible 2- and 4- part proximal humeral fractures; avascular necrosis of the humeral head; or, other difficult clinical management problems where arthrodesis or resectional arthroplasty is not acceptable. The glenoid components are designed for cemented fixation in the joint and must only be used with appropriate bone cement.
Performance Data	Mechanical testing (i.e. Rocking horse testing per ASTM F2028) was performed to demonstrate that the proposed device meets the standards requirements. Bacterial Endotoxin test was conducted in accordance with ANSI/AAMI ST72:2011/(R)2016, USP <161>, USP <85>, EP 2.6.14 to demonstrate that the proposed device meets pyrogen limit specifications. MRI testing was conducted in accordance with FDA guidance Testing and Labeling Medical Devices for Safety in the Magnetic Resonance (MR) Environment and ASTM F2182.
Conclusion	The Augmented VaultLock Glenoid is substantially equivalent to the predicate device in which the basic design features and intended use are the same. The mechanical testing data demonstrates that the proposed device performance is equivalent to the predicate device for the desired indications. Any differences between the proposed device and the predicate device are considered minor and do not raise questions regarding safety or effectiveness. Based on the indications for use, technological characteristics, and the summary of data submitted, Arthrex Inc. has determined that the proposed device is substantially equivalent to the currently marketed predicate device.